


United States Department of Commerce  
National Institute of Standards and Technology



---

**Certificate of Accreditation to ISO/IEC 17025:2017**

---

**NVLAP LAB CODE: 100518-0**

**Landauer, Inc.**  
Glenwood, IL

*is accredited by the National Voluntary Laboratory Accreditation Program for specific services,  
listed on the Scope of Accreditation, for:*


**Ionizing Radiation Dosimetry**

*This laboratory is accredited in accordance with the recognized International Standard ISO/IEC 17025:2017.  
This accreditation demonstrates technical competence for a defined scope and the operation of a laboratory quality  
management system (refer to joint ISO-ILAC-IAF Communique dated January 2009).*

---

2021-01-01 through 2021-12-31

*Effective Dates*



---

*For the National Voluntary Laboratory Accreditation Program*

---

**SCOPE OF ACCREDITATION TO ISO/IEC 17025:2017**

**Landauer, Inc.**  
2 Science Road  
Glenwood, IL 60425  
Mr. Christopher N. Passmore, CHP  
Phone: 708-441-8455 Fax: 708-755-7035  
E-Mail: [cpassmore@landauerinc.com](mailto:cpassmore@landauerinc.com)  
URL: <http://www.landauer.com>

**IONIZING RADIATION DOSIMETRY**

**NVLAP LAB CODE 100518-0**

***Scope of Accreditation:***

This facility has been evaluated and deemed competent to process the whole body radiation dosimeters and extremity dosimeters listed below by employing the following readers/processes:

- (1) Landauer InLight manual, 200, 500, 3000 and microStar Optically Stimulated Luminescence (OSL) readers; (2) LDR Custom Laser Heater TLD Reader; (3) Harshaw 3500 single-chip reader; (4) Landauer Automatic and Manual Luxel/Luxel+ Pulsed Optically Stimulated Luminescence (POSL) reader; (5) Landauer Manual Luxel/Luxel+ Light Emitting Diode (LED) OSL accident reader; (6) Neutron Auto CR-39 reader RadOsys; (7) microscopes; and (8) RadLight OSL manual and automatic readers.

---

**WHOLE BODY**

This facility is accredited to process the following dosimeters by demonstration of compliance with ANSI HPS N13.11 through testing.

**OSL Dosimeters:**

A4 - [Beta/photon] - Luxel+ Pa (w/permanent clip), process 5 and 6 for ANSI N13.11-2009 categories IA, IIA, IIIA, and IVAA.

A5 - [Beta/photon/thermal & fast neutrons] - Luxel+ Ta (w/permanent clip), process 5, 6, 7, and 8 for ANSI N13.11-2009 categories IA, IIA, IIIA, IVAA and VCA.

A7 - [Beta/photon] - InLight LDR Model 2-L02N (w/permanent clip), process 1 for ANSI N13.11-2009 categories IA, IIA, IIIA, and IVAA.



---

2021-01-01 through 2021-12-31

*Effective dates*

*For the National Voluntary Laboratory Accreditation Program*

**IONIZING RADIATION DOSIMETRY**

**NVLAP LAB CODE 100518-0**

A9 - [Beta/photon/thermal & fast neutrons] - InLight LDR Model 2T-L02T (Polyethylene and Boron Radiator CR39) (w/permanent clip), process 1, 7, and 8 for ANSI N13.11-2009 categories IA, IIA, IIIA, IVAA, and VCA.

B3 - [Beta/photon/fast neutrons] - Luxel+ Ja (w/permanent clip), process 5, 6, 7, and 8 for ANSI N13.11-2009 categories IA, IIA, IIIA, IVAA, and VCA.

B6 - [Photon] - Luxel+ Pa Escort, process 5 and 6 for ANSI N13.11-2009 category IA.

C1 - [Beta/photon/ neutron] - InLight LDR Model 2 OSLN-L11N, process 1 for ANSI N13.11-2009 categories IA, IIA, IIIA, IVAA and VCA.

C2 - [Beta/photon/thermal & fast neutrons] - InLight LDR Model 2T OSLN-L11TN, process 1, 7, and 8 for ANSI N13.11-2009 categories IA, IIA, IIIA, IVAA, and VCA.

C3 - [Photon] - RadWatch Model 1 in RadWatch holder, process 9 for ANSI N13.11-2009 categories IA and IIC.

---

**EXTREMITY**

This facility is accredited to process the following dosimeters by demonstration of compliance with ANSI HPS N13.32 through testing.

C5 - [Beta/photon] - S Ring (TLD 100-chip) (Finger), process 2 and 4 for ANSI N13.32-2008 categories IB, IC, IIB, IIC, IID, IIIB, IIIC, IIID, and IVBB.

OSL Badges:

B2 - [Beta/photon] - Luxel+ Pa Wrist (Wrist), process 5 and 6 for ANSI N13.32-2008 categories IA, IIA, IIIA, and IVAA.

---

2021-01-01 through 2021-12-31

*Effective dates*



---

*For the National Voluntary Laboratory Accreditation Program*